

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627105.

27/05/2024

MRL/REF/RDCH/ 210720

Date: 23/05/2024

From

To

Name : Dr. T. K Divakar MDS.

The Chairman, RDCH

Designation : Reader ,

Dept. : OMFS PG.

Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

SL. No.	Nature of Work	Place of Work
	Spittoon damage CHAIR NO: 3. (water leakage)	Clinical Area
	CHAIR NO: 4 Spittoon damage water leakage.	Clinical Area.
	Name and Signature :	Completion / Non Completion Report
Staff who submits the MRL	P. K. Divakar	
Recommendation with estimated cost of the work (attach details)		
Chair No: 3 drainage Hose full set replaced and serviced.		Work Completed the repair/ maintenance work as required in MRL on 27/05/2024. It works smoothly.
Chair No: 4 spittoon loose rectified		
Technical Staff	Co-ordinator	Staff
		Co-ordinator

Approved/Rejected with remarks

A. B. S. 25/5/2024

CFO/CAO



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627105.

MRL/REF/RDCH/A/0610

Date: 12/04/2024

From

To

Name : Dr. T.K. Divakar MDS

The Chairman, RDCH

Designation : Reader

Dept. : OMFS PG


Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

SL. No.	Nature of Work	Place of Work
1	Printer color print not working	HED Room
2	Loosening of cord - micromotor ①	Clinical Area
3	Department handpiece - not working	
Name and Signature :		Completion / Non Completion Report
Staff who submits the MRL	T.K. Divakar	
Recommendation with estimated cost of the work (attach details)		Work Completed the repair/ maintenance work as required in MRL on 25.04.24. It works smoothly.
1. CHAIR NO 1 Micro motor Female Adaptor Replaced.		
2. Department straight hand Piece Full service.		Dr. S. Reddy
Technical Staff	Co-ordinator	
		Co-ordinator

Approved/Rejected with remarks

CFO/CAO



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627105.

MRL/REF/RDCH/ A/01032

Date: 20/08/2024.

From

To

Name : Dr. Divakar MD


The Chairman, RDCH

Designation : Reader

Dept. : OMFS PGI.

Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

SL. No.	Nature of Work	Place of Work
1	Straight handpiece Service - 3 no.	OMFS - PGI
2	Handpiece cord not working - 3 no	OMFS - PGI.
3.	Bending B.P Apparatus Not working	OMFS - PGI. - <u>Urgently required</u>
Name and Signature :		Completion / Non Completion Report
Staff who submits the MRL	P.K. Divakar	 (27/9/24, Dr. Divakar)
Recommendation with estimated cost of the work (attach details)		Work Completed the repair/ maintenance work as required in MRL on 27/09/2024. It works smoothly.
Straight handpiece Serviced - 3 Nos Hand piece - 1 Carbon brush, lock system, Mains cord - 2 set replaced. 22/08/2024 BP Apparatus serviced on 27/09/2024 C Pulp and Valve replaced?		
Technical Staff	Co-ordinator	Staff Co-ordinator

Approved/Rejected with remarks

G. Balaji
20/8/2024

CFO/CAO



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627105.

MRL/REF/RDCH/ A/0673

Date: 06/05/2024.

From

To

Name : Dr. Sindhuadevi MDS

The Chairman, RDCH

Designation : Senior Lect

Dept. : OMFS PG

Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

SL. No.	Nature of Work	Place of Work
1.	Ac is not working in clinical area	clinical area.
2.	Two bulbs not working in clinical area	clinical area.
	Name and Signature	Completion / Non Completion Report
Staff who submits the MRL	Dr. S. Sindhuadevi	
Recommendation with estimated cost of the work (attach details)		
		Work. Completed the repair/ maintenance work as required in MRL on . It works smoothly.
Technical Staff	Co-ordinator	Staff
		Co-ordinator

Approved/Rejected with remarks

G. Balaji
6/5/24
CFO/CAO



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627105.

MRL/REF/RDCH/ A/01032

Date: 20/08/2024.

From

To

Name : Dr. Divakar MD


The Chairman, RDCH

Designation : Reader

Dept. : DMFS Pgt.

Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

SL. No.	Nature of Work	Place of Work
1	Straight handpiece Service - 3 no.	DMFS - Pgt.
2	Handpiece cord not working - 3 no	DMFS - Pgt.
3	Bending B.R Apparatus Not working	DMFS - Pgt. - <u>Urgently require</u>
Name and Signature :		Completion / Non Completion Report
Staff who submits the MRL	P.K. Divakar	 (27/09/24. Dr. Divakar)
Recommendation with estimated cost of the work (attach details)		MRL. Completed the repair/ maintenance work as required in MRL on 27/09/2024. It works smoothly.
Straight handpiece Serviced - 3 Nos Hand piece 1 Carbon brush, Lock system. Mains cord - 2 set replaced. 22/08/2024 BP Apparatus Serviced on 27/09/2024 CPulp and Valve replaced?		
Technical Staff	Co-ordinator	Staff Co-ordinator

Approved/Rejected with remarks

G. Balaji
20/08/2024

CFO/CAO



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627 105.

MRL/REF/RDCH/A-1201

Date: 18/10/2021

From

To

Name : Dr. Divakar MDS


The Chairman, RDCH

Designation : Reader

Dept. : OMFS PTG

Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

Sl. No.	Nature of Work	Place of Work
1	5 Ceiling light not working.	Clinical area.
	Name and Signature	Completion / Non Completion Report
Staff who submits the MRL	Divakar	
Recommendation with estimated cost of the work (attach details)		Work Completed the repair / maintenance work as required in MRL on smoothly. It works
Technical Staff	Co-ordinator	Staff
		Co-ordinator

Approved/Rejected with remarks

G. Balaji 18/10/2021

C.E.O.



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627 105.

MRL/REF/RDCH/A-1200

Date: 18/10/2020

From

To

Name : Dr. Divakar MDS


The Chairman, RDCH

Designation : Reader

Dept. : OMFS OT

Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

Sl. No.	Nature of Work	Place of Work
1	AC not working (Malfunctioning)	ICU
2	AC not working	OT
3	Hand Shower not working in rest room.	OT (Bathroom)
Name and Signature		Completion / Non Completion Report
Staff who submits the MRL	Divakar	
Recommendation with estimated cost of the work (attach details)	 Completed the repair / maintenance work as required in MRL on . It works smoothly.
Technical Staff	Co-ordinator	Staff
		Co-ordinator

Approved/Rejected with remarks

G. Balaji 18/10/2020

C.E.O.



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627 105.

MRL/REF/RDCH /A-1216

Date: 18/10/2024.

From

To

Name : Dr. Divakar MDS




The Chairman, RDCH

Designation : Reader

Dept. : DMFS PG

Repairs & Maintenance Work-Type


I report the following repair and maintenance work for your immediate action.

Sl. No.	Nature of Work	Place of Work
	Foot Pedal Not working in chair 3 & screw fixation.	} clinical area.
	Chair 1 - X-ray light not working	
	Chair 2 - Slow Speed in Micromotor.	
	camera Not working (Department)	DMFS PG
Name and Signature		Completion / Non Completion Report
Staff who submits the MRL		<p>Completed the repair / maintenance work as required in MRL on 28/10/2024. It works smoothly.</p>
Recommendation with estimated cost of the work (attach details)		
<p>Chair : 3 foot pedal replaced.</p> <p>Chair 1, 2 X-ray Viewer replaced.</p>		
<p> Technical Staff</p>		<p> Co-ordinator</p>

Approved/Rejected with remarks

G. Balaji 24/10/24

C.E.O.



CHAIRMAN

RAJAS DENTAL COLLEGE & HOSPITAL

KAVALKINARU JN. - 627 105. Received on 22/11/2024

MRL/REF/RDCH /A-1275

Date: 18/11/2024

From

To

Name : Dr. Divakar MDS

The Chairman, RDCH

Designation : Reader

Dept. : DMFS PG

Repairs & Maintenance Work-Type

I report the following repair and maintenance work for your immediate action.

Sl. No.	Nature of Work	Place of Work
	Micromotor not working - chair no ③	clinical area
	switch broken in op Chair. - Hand Engine not working ④ chair No. 2	"
	Slow Speed Micromotor in chair ②	"
	Name and Signature :	Completion / Non Completion Report
Staff who submits the MRL	T.K. Divakar	Dr. T.K. Divakar
Recommendation with estimated cost of the work (attach details)		Work Completed the repair / maintenance work as required in MRL on 14/12/2024. It works smoothly.
switch replaced in op chair No: 2 foot pedal wire cut rectified. Completed on 25/11/2024		
Need Micromotor Bearing (2 Micromotor - 4 Bearing - 2 set) Carbon brush replaced On 14/12/2024		
Technical Staff	Co-ordinator	Staff Co-ordinator

Approved/Rejected with remarks

A. Balaji 18/11/24

C.E.O.

CHAIRMAN